



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 3540			
EDTA TITRANT 0.1 MOLAR (0.2N)			
Test	Target/UOM	Range	Result
Actual Molarity	0.1000 MOL/L	0.0995 - 0.1005	0.1003
Sample ID			
Tested	07/31/20		
Tester	HSID		
Recorded	07/31/20		
QCSP4:STANDARDIZATION EDTA SOLUTIONS	YES QCSP4		YES
Sample ID			
Tested	07/31/20		
Tester	HSID		
Recorded	07/31/20		
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	07/31/20		
Tester	HSID		
Recorded	07/31/20		
FORMULA WEIGHT	372.24 F.W.	372.24 - 372.24	372.24
Sample ID			
Tested	07/31/20		
Tester	HSID		
Recorded	07/31/20		
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested	07/31/20		
Tester	HSID		
Recorded	07/31/20		
Actual Density	1.01671 GM/ML	1.01605 - 1.01736	1.01644
Sample ID			
Tested	07/31/20		
Tester	HSID		
Recorded	07/31/20		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M85,T27,M55,M81,BID02185,TI DS31610
Sample ID			
Tested	07/31/20		
Tester	HSID		
Recorded	07/31/20		
Traceable to N.I.S.T. SRM 915	YES YES		YES
Sample ID			
Tested	07/31/20		
Tester	HSID		
Recorded	07/31/20		



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

EXPIRATION DATE	MM/DD/YY MM/DD/YYYY	02/28/22
Sample ID		
Tested	07/31/20	
Tester	HSID	
Recorded	07/31/20	
EXPIRATION FOR FANN PRODUCTS	MM/DD/YY EXP	08/30/22
Sample ID		
Tested	07/31/20	
Tester	HSID	
Recorded	07/31/20	
SHELF LIFE OF PRODUCT	SHELF LIFE TIME	18 MONTHS
Sample ID		
Tested	07/31/20	
Tester	HSID	
Recorded	07/31/20	
SHELF LIFE FOR FANN PRODUCTS	2 YEARS SHELF LIFE	2 YEARS
Sample ID		
Tested	07/31/20	
Tester	HSID	
Recorded	07/31/20	
Lot #	0073082	
Made	07/31/20	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.


Quality Manager