



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| Packaged Product: HON009F | | | |
|--|-------------------------|-------------|-------------|
| QC Calibration Std. 1.5 mg/L Cl Solution | | | |
| Test | Target/UOM | Range | Result |
| PREPARED TO FORMULATION ON FILE | YES | | YES |
| Sample ID | YES | | |
| Tested 07/10/23 | | | |
| Tester MMAT | | | |
| Recorded 07/10/23 | | | |
| PPM CHLORIDE | 1.50 | 1.48 - 1.52 | 1.50 |
| Sample ID | PPM | | |
| Tested 07/10/23 | | | |
| Tester MMAT | | | |
| Recorded 07/10/23 | | | |
| APPEARANCE AND COLOR | CLEAR LIQUID APPEARANCE | | PASS |
| Sample ID | | | |
| Tested 07/10/23 | | | |
| Tester MMAT | | | |
| Recorded 07/10/23 | | | |
| INSTRUMENTS USED DURING PREPARATION | INSTRUMENT | | P384,VF2-A3 |
| Sample ID | | | |
| Tested 07/10/23 | | | |
| Tester MMAT | | | |
| Recorded 07/10/23 | | | |
| Lot # | 30710121 | | |
| Made | 07/10/23 | | |

Quality Manager