



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 9060			
SULFURIC ACID 0.25 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	0.2500	0.2495 - 0.2505	0.2499
	EQ/L		
Sample ID			
Tested	08/03/23		
Tester	JENC		
Recorded	08/03/23		
QCSP5: STANDARDIZATION OF AQUEOUS ACIDS	YES		YES
	METHOD		
Sample ID			
Tested	08/03/23		
Tester	JENC		
Recorded	08/03/23		
N.I.S.T. Traceable to SRM 723	Yes		YES
	YES		
Sample ID			
Tested	08/03/23		
Tester	JENC		
Recorded	08/03/23		
PREPARED TO FORMULATION ON FILE	YES		YES
	YES		
Sample ID			
Tested	08/03/23		
Tester	JENC		
Recorded	08/03/23		
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
	YES		
Sample ID			
Tested	08/03/23		
Tester	JENC		
Recorded	08/03/23		
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		08/03/23
	MM/DD/YYYY		
Sample ID			
Tested	08/03/23		
Tester	JENC		
Recorded	08/03/23		
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0	10.0 - 35.0	22.6
	DEG C		
Sample ID			
Tested	08/03/23		
Tester	JENC		
Recorded	08/03/23		
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		08/03/23
	MM/DD/YYYY		
Sample ID			
Tested	08/03/23		
Tester	JENC		
Recorded	08/03/23		



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	22.6
DEG C	DEG C		
Sample ID			
Tested	08/03/23		
Tester	JENC		
Recorded	08/03/23		
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested	08/03/23		
Tester	JENC		
Recorded	08/03/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		GC,T-36,M55,BID02259,PIDM9
			1
Sample ID			
Tested	08/03/23		
Tester	JENC		
Recorded	08/03/23		
EXPIRATION DATE	MM/DD/YY		08/30/25
	MM/DD/YYYY		
Sample ID			
Tested	08/03/23		
Tester	JENC		
Recorded	08/03/23		
Actual Density	1.00653	1.00646 - 1.00660	1.00652
	GM/ML		
Sample ID			
Tested	08/03/23		
Tester	JENC		
Recorded	08/03/23		
Lot #	3080266		
Made	08/03/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager