



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: OXY204F			
LACHAT CALIBRATION STANDARD LEVEL 4			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 08/14/23			
Tester AFRE			
Recorded 08/14/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE APPEARANCE		PASS
Sample ID			
Tested 08/14/23			
Tester AFRE			
Recorded 08/14/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P125,P122,VF6-W8
Sample ID			
Tested 08/14/23			
Tester AFRE			
Recorded 08/14/23			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		08/30/24
Sample ID			
Tested 08/14/23			
Tester AFRE			
Recorded 08/14/23			
PPM KIO3-NAOCL	30.0 PPM	29.9 - 30.1	30.0
Sample ID			
Tested 08/14/23			
Tester AFRE			
Recorded 08/14/23			
CALCULATED GPL SODIUM HYDROXIDE	125.00 GPL	124.95 - 125.05	125.00
Sample ID			
Tested 08/14/23			
Tester AFRE			
Recorded 08/14/23			
CALCULATED GPL SODIUM CHLORIDE	155.00 GPL	154.98 - 155.02	155.00
Sample ID			
Tested 08/14/23			
Tester AFRE			
Recorded 08/14/23			
CALCULATED GPL SODIUM CHLORATE	0.8000 GPL	0.7995 - 0.8005	0.8000
Sample ID			
Tested 08/14/23			
Tester AFRE			
Recorded 08/14/23			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

Actual Density	1.21025	1.20705 - 1.21345	1.21184
	GM/ML		
Sample ID			
Tested	08/14/23		
Tester	AFRE		
Recorded	08/14/23		
Lot #	3081437		
Made	08/14/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager