



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 9563			
TOTAL ORGANIC CARBON STANDARD 5.0 mg/L (ppm)			
Test	Target/UOM	Range	Result
SHELF LIFE OF PRODUCT			
	SHELF LIFE		2 MONTHS
	TIME		
Sample ID			
Tested	08/14/23		
Tester	DTRA		
Recorded	08/14/23		
PREPARED TO FORMULATION ON FILE			
	YES		YES
	YES		
Sample ID			
Tested	08/14/23		
Tester	DTRA		
Recorded	08/14/23		
PRESERVED WITH HYDROCHLORIC ACID			
	YES		YES
	YES		
Sample ID			
Tested	08/14/23		
Tester	DTRA		
Recorded	08/14/23		
PPM TOC			
	5.00	4.70 - 5.30	5.14
	PPM		
Sample ID			
Tested	08/14/23		
Tester	DTRA		
Recorded	08/14/23		
APPEARANCE AND COLOR			
	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested	08/14/23		
Tester	DTRA		
Recorded	08/14/23		
INSTRUMENTS USED DURING PREPARATION			
	INSTRUMENT		P78,GC,VF1-Z39
Sample ID			
Tested	08/14/23		
Tester	DTRA		
Recorded	08/14/23		
EXPIRATION DATE			
	MM/DD/YY		10/19/23
	MM/DD/YYYY		
Sample ID			
Tested	08/14/23		
Tester	DTRA		
Recorded	08/14/23		
Lot #	3081450		
Made	08/14/23		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 08/14/23 at 3:09 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager