



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> AXI035F			
Working Solution #2S (1L)			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 08/21/23			
Tester AROS			
Recorded 08/21/23			
<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 08/21/23			
Tester AROS			
Recorded 08/21/23			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M100,P54,VF1-Z20
Sample ID			
Tested 08/21/23			
Tester AROS			
Recorded 08/21/23			
<b>SHELF LIFE OF PRODUCT</b>	1 YEAR TIME		1 YEAR
Sample ID			
Tested 08/21/23			
Tester AROS			
Recorded 08/21/23			
<b>EXPIRATION DATE</b>	MM/DD/YYYY		08/30/24
Sample ID	MM/DD/YYYY		
Tested 08/21/23			
Tester AROS			
Recorded 08/21/23			
<b>CALCULATED GPL SODIUM CHLORIDE</b>	100.0 GPL	99.9 - 100.1	100.0
Sample ID			
Tested 08/21/23			
Tester AROS			
Recorded 08/21/23			
<b>GPL SODIUM CHLORATE</b>	2.000 GPL	1.980 - 2.020	2.000
Sample ID			
Tested 08/21/23			
Tester AROS			
Recorded 08/21/23			
<b>CALCULATED MG/L SODIUM HYPOCHLORITE</b>	40.0 MG/L (PPM)	39.8 - 40.2	40.0
Sample ID			
Tested 08/21/23			
Tester AROS			
Recorded 08/21/23			



Aqua Solutions, Inc.  
 Ph: 281-479-2569  
 Fax: 281-479-2790  
 P.O. Box 70  
 Deer Park, TX 77536  
 Website: [www.aquasolutions.org](http://www.aquasolutions.org)

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**


Customer PO # :

<b>CALCULATED GPL SODIUM HYDROXIDE</b>	200.0	199.9 - 200.1	200.0
	GPL		
Sample ID			
Tested	08/21/23		
Tester	AROS		
Recorded	08/21/23		
<b>Lot #</b>	3082177		
<b>Made</b>	08/21/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

  
**Quality Manager**