



Aqua Solutions, Inc.  
 Ph: 281-479-2569  
 Fax: 281-479-2790  
 P.O. Box 70  
 Deer Park, TX 77536  
 Website: [www.aquasolutions.org](http://www.aquasolutions.org)

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> DC994F			
35°C HYDROXYL REAGENT			
<b>Test</b>	<b>Target/UOM</b>	<b>Range</b>	<b>Result</b>
<b>PREPARED TO FORMULATION ON</b>			
	YES		YES
<b>FILE</b>			
	YES		
Sample ID			
Tested	09/14/23		
Tester	MMAT		
Recorded	09/14/23		
<b>APPEARANCE AND COLOR</b>			
	CLEAR YELLOW TO TAN TO		PASS
	BROWN LIQUI		
	APPEARANCE		
Sample ID			
Tested	09/14/23		
Tester	MMAT		
Recorded	09/14/23		
<b>INSTRUMENTS USED DURING PREPARATION</b>			
	INSTRUMENT		GC,EF2-SD,M100
Sample ID			
Tested	09/14/23		
Tester	MMAT		
Recorded	09/14/23		
<b>SHELF LIFE OF PRODUCT</b>			
	9 MONTHS		9 MONTHS
	TIME		
Sample ID			
Tested	09/14/23		
Tester	MMAT		
Recorded	09/14/23		
<b>EXPIRATION DATE</b>			
	MM/DD/YYYY		06/30/2024
	MM/DD/YYYY		
Sample ID			
Tested	09/14/23		
Tester	MMAT		
Recorded	09/14/23		
<b>Lot #</b>	3091185		
<b>Made</b>	09/14/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.