



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Test	Target/UOM	Range	Result
Packaged Product: 9764-4L			
	WATER, DEIONIZED REAGENT GRADE ASTM TYPE III (4L)		
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 09/18/23			
Tester TBUR			
Recorded 09/18/23			
PREPARED TO ASTM D 1193	PASS		PASS
Sample ID	PASS		
Tested 09/18/23			
Tester TBUR			
Recorded 09/18/23			
MEETS ASTM D 1193 TYPE III SPECIFICATION	PASS		PASS
Sample ID	PASS		
Tested 09/18/23			
Tester TBUR			
Recorded 09/18/23			
PPB toc	<200 PPB		<200 PPB
Sample ID	PPB		
Tested 09/18/23			
Tester TBUR			
Recorded 09/18/23			
PPB Sodium	<10 PPB		<10 PPB
Sample ID	PPB		
Tested 09/18/23			
Tester TBUR			
Recorded 09/18/23			
PPB Chloride	<10 PPB		<10 PPB
Sample ID	PPB		
Tested 09/18/23			
Tester TBUR			
Recorded 09/18/23			
PPB Silica	500 PPB OR LESS		500 PPB OR LESS
Sample ID	PPB		
Tested 09/18/23			
Tester TBUR			
Recorded 09/18/23			
CONDUCTIVITY	<0.25 UMHOS		<0.25 UMHOS
Sample ID	US/CM		
Tested 09/18/23			
Tester TBUR			
Recorded 09/18/23			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

MEGOHM-CMIN 1	MIN 4.0 MEGOHM-CM MIN 1	MIN 4.0 MEGOHM-CM
Sample ID		
Tested 09/18/23		
Tester TBUR		
Recorded 09/18/23		
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE	CLEAR LIQUID
Sample ID		
Tested 09/18/23		
Tester TBUR		
Recorded 09/18/23		
SHELF LIFE OF PRODUCT	1 YEAR TIME	1 YEAR
Sample ID		
Tested 09/18/23		
Tester TBUR		
Recorded 09/18/23		
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY	07/30/2024
Sample ID		
Tested 09/18/23		
Tester TBUR		
Recorded 09/18/23		
Lot # 3072499		
Made 07/25/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.


Quality Manager