



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: HAC440F			
Alkalinity Cal Soln. 1,250 mg/L CaCo3 - EZ5006 Hardness			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested 10/04/23			
Tester MMAT			
Recorded 10/04/23			
PPM AS CaCO3	1250 PPM	1240 - 1260	1250
Sample ID			
Tested 10/04/23			
Tester MMAT			
Recorded 10/04/23			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 10/04/23			
Tester MMAT			
Recorded 10/04/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M115,VF4-N27
Sample ID			
Tested 10/04/23			
Tester MMAT			
Recorded 10/04/23			
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY		02/28/2024
Sample ID			
Tested 10/04/23			
Tester MMAT			
Recorded 10/04/23			
Lot #	3100433		
Made	10/04/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.