



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: ERL083F			
BROMIDE-BROMATE 0.500 NORMAL-1363			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested	10/05/23		
Tester	MHUS		
Recorded	10/05/23		
Actual Normality	0.5000 EQ/L	0.4800 - 0.5200	0.5001
Sample ID			
Tested	10/05/23		
Tester	MHUS		
Recorded	10/05/23		
QCSP11:STANDARDIZATION OF BROMIDE-BROMAT	YES QCSP11		YES
Sample ID			
Tested	10/05/23		
Tester	MHUS		
Recorded	10/05/23		
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested	10/05/23		
Tester	MHUS		
Recorded	10/05/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M55,BID3902,M91,P165,TID41 123
Sample ID			
Tested	10/05/23		
Tester	MHUS		
Recorded	10/05/23		
Actual Density	1.04444 GM/ML	1.04354 - 1.04535	1.04382
Sample ID			
Tested	10/05/23		
Tester	MHUS		
Recorded	10/05/23		
SHELF LIFE OF PRODUCT	1 MONTH TIME		1 MONTH
Sample ID			
Tested	10/05/23		
Tester	MHUS		
Recorded	10/05/23		
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY		11/15/2023
Sample ID			
Tested	10/05/23		
Tester	MHUS		
Recorded	10/05/23		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 10/05/23 at 1:50 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Lot #	3091819
Made	10/05/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager