



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: FIS101F			
Alkalinity Standard 90 ppm Certified			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 10/18/23			
Tester AFRE			
Recorded 10/18/23			
ALKALINITY AS CaCO3	90.0	85.0 - 95.0	90.6
	PPM		
Sample ID			
Tested 10/18/23			
Tester AFRE			
Recorded 10/18/23			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 10/18/23			
Tester AFRE			
Recorded 10/18/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M91,VF1-Z28,BID-00488,PID-1 27
Sample ID			
Tested 10/18/23			
Tester AFRE			
Recorded 10/18/23			
EXPIRATION DATE	MM/DD/YYYY		02/28/24
Sample ID			
Tested 10/18/23			
Tester AFRE			
Recorded 10/18/23			
Lot #	3101255		
Made	10/18/23		

Quality Manager