



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 4099			
Hydrochloric Acid 0.025 Normal NIST Traceable solution			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID	YES		
Tested 10/18/23			
Tester ENOR			
Recorded 10/18/23			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID	YES		
Tested 10/18/23			
Tester ENOR			
Recorded 10/18/23			
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY		10/18/2023
Sample ID	MM/DD/YYYY		
Tested 10/18/23			
Tester ENOR			
Recorded 10/18/23			
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0	10.0 - 35.0	20.5
Sample ID	DEG C		
Tested 10/18/23			
Tester ENOR			
Recorded 10/18/23			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY		10/18/2023
Sample ID	MM/DD/YYYY		
Tested 10/18/23			
Tester ENOR			
Recorded 10/18/23			
<b>AMBIENT TEST TEMPERATURE DEG C</b>	20.0	15.0 - 30.0	20.0
Sample ID	DEG C		
Tested 10/18/23			
Tester ENOR			
Recorded 10/18/23			
<b>NORMALITY TO 4 DEC PLACES</b>	0.0250	0.0245 - 0.0255	0.0245
Sample ID	EQUIV/LITER		
Tested 10/18/23			
Tester ENOR			
Recorded 10/18/23			
<b>QCSP5: STANDARDIZATION OF AQUEOUS ACIDS</b>	YES		YES
Sample ID	METHOD		
Tested 10/18/23			
Tester ENOR			
Recorded 10/18/23			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>N.I.S.T. TRACEABLE TO SRM 723</b>	YES YES	YES
Sample ID		
Tested	10/18/23	
Tester	ENOR	
Recorded	10/18/23	
<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID APPEARANCE	PASS
Sample ID		
Tested	10/18/23	
Tester	ENOR	
Recorded	10/18/23	
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT	VF5-L9,T-243,M55,BID69690,P IDM91
Sample ID		
Tested	10/18/23	
Tester	ENOR	
Recorded	10/18/23	
<b>Lot #</b>	3101205	
<b>Made</b>	10/18/23	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**