



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 4835			
IODINE MONOCHLORIDE WIJS SOLUTION AOAC FOR IODINE ABSORPTION			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 10/25/23			
Tester AFRE			
Recorded 10/25/23			
TOTAL HALOGEN RATION (I/CL)	1.10	1.00 - 1.20	1.05
	NUMBER		
Sample ID			
Tested 10/25/23			
Tester AFRE			
Recorded 10/25/23			
APPEARANCE AND COLOR	BROWN LIQUID APPEARANCE		PASS
Sample ID			
Tested 10/25/23			
Tester AFRE			
Recorded 10/25/23			
PREPARED TO AOAC METHOD	YES		YES
	YES		
Sample ID			
Tested 10/25/23			
Tester AFRE			
Recorded 10/25/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M103,T-20,M55,BID03828,PID P67
Sample ID			
Tested 10/25/23			
Tester AFRE			
Recorded 10/25/23			
EXPIRATION DATE	MM/DD/YY		04/30/25
	MM/DD/YYYY		
Sample ID			
Tested 10/25/23			
Tester AFRE			
Recorded 10/25/23			
NORMALITY OF PART A	0.2025	0.2015 - 0.2035	0.2029
	EQUIV/LITER		
Sample ID			
Tested 10/25/23			
Tester AFRE			
Recorded 10/25/23			
NORMALITY OF PART B	0.6275	0.6050 - 0.6500	0.6261
	EQUIV/LITER		
Sample ID			
Tested 10/25/23			
Tester AFRE			
Recorded 10/25/23			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

EXPIRATION DATE	MM/DD/YYYY	04/30/2025
	MM/DD/YYYY	
Sample ID		
Tested	10/25/23	
Tester	AFRE	
Recorded	10/25/23	
Lot #	3091947	
Made	10/25/23	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager