



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: ND691F			
Mixed ICP 0.03 PPM Pb, Na, K, Ni, Cr, Ca, Cu, and Fe Standard			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 12/06/23			
Tester MMAT			
Recorded 12/06/23			
PPM OF LEAD TO 3 DECIMAL PLACES	0.030	0.027 - 0.033	0.030
Sample ID	PPM		
Tested 12/06/23			
Tester MMAT			
Recorded 12/06/23			
PPM Sodium(Na)	0.030	0.027 - 0.033	0.030
Sample ID	1.000		
Tested 12/06/23			
Tester MMAT			
Recorded 12/06/23			
PPM Potassium(K)	0.030	0.027 - 0.033	0.030
Sample ID	1.000		
Tested 12/06/23			
Tester MMAT			
Recorded 12/06/23			
PPM Nickel(Ni)	0.030	0.027 - 0.033	0.030
Sample ID	1.000		
Tested 12/06/23			
Tester MMAT			
Recorded 12/06/23			
PPM Chromium(Cr)	0.030	0.027 - 0.033	0.030
Sample ID	1.000		
Tested 12/06/23			
Tester MMAT			
Recorded 12/06/23			
PPM Calcium(Ca)	0.030	0.027 - 0.033	0.030
Sample ID	1.000		
Tested 12/06/23			
Tester MMAT			
Recorded 12/06/23			
PPM Copper(Cu)	0.030	0.027 - 0.033	0.030
Sample ID	1.000		
Tested 12/06/23			
Tester MMAT			
Recorded 12/06/23			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

Calculated ppm (mg/L) Iron	0.030	0.027 - 0.033	0.030
	MG/L		
Sample ID			
Tested	12/06/23		
Tester	MMAT		
Recorded	12/06/23		
APPEARANCE AND COLOR	PASS		PASS
	APPEARANCE		
Sample ID			
Tested	12/06/23		
Tester	MMAT		
Recorded	12/06/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P103,GC,VF500-X30
Sample ID			
Tested	12/06/23		
Tester	MMAT		
Recorded	12/06/23		
EXPIRATION DATE	MM/DD/YYYY		06/30/2024
	MM/DD/YYYY		
Sample ID			
Tested	12/06/23		
Tester	MMAT		
Recorded	12/06/23		
Lot #	3120501		
Made	12/06/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager