



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 4270			
HYDROCHLORIC ACID 0.773 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 12/04/23			
Tester AFRE			
Recorded 12/04/23			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 12/04/23			
Tester AFRE			
Recorded 12/04/23			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		12/04/2023
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 12/04/23			
Tester AFRE			
Recorded 12/04/23			
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	20.0
SAMPLING DEGC	DEG C		
Sample ID			
Tested 12/04/23			
Tester AFRE			
Recorded 12/04/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		12/04/2023
	MM/DD/YYYY		
Sample ID			
Tested 12/04/23			
Tester AFRE			
Recorded 12/04/23			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	20.0
DEG C	DEG C		
Sample ID			
Tested 12/04/23			
Tester AFRE			
Recorded 12/04/23			
NORMALITY TO 4 DEC PLACES	0.7730	0.7725 - 0.7735	0.7734
	EQUIV/LITER		
Sample ID			
Tested 12/04/23			
Tester AFRE			
Recorded 12/04/23			
N.I.S.T. Traceable to SRM 723 THAM	YES		YES
	YES		
Sample ID			
Tested 12/04/23			
Tester AFRE			
Recorded 12/04/23			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE	PASS
Sample ID		
Tested	12/04/23	
Tester	AFRE	
Recorded	12/04/23	
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	VF6-W7,GC,T-243,M55,BID004 88,PIDM9
Sample ID		
Tested	12/04/23	
Tester	AFRE	
Recorded	12/04/23	
Lot #	3113023	
Made	12/04/23	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager