



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 7286			
POTASSIUM HYDROXIDE 2.0 NORMAL			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 12/05/23			
Tester AFRE			
Recorded 12/05/23			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID			
Tested 12/05/23			
Tester AFRE			
Recorded 12/05/23			
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY		12/05/2023
Sample ID			
Tested 12/05/23			
Tester AFRE			
Recorded 12/05/23			
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0	10.0 - 35.0	20.0
Sample ID			
Tested 12/05/23			
Tester AFRE			
Recorded 12/05/23			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY		12/05/2023
Sample ID			
Tested 12/05/23			
Tester AFRE			
Recorded 12/05/23			
<b>AMBIENT TEST TEMPERATURE DEGC</b>	20.0	15.0 - 30.0	20.0
Sample ID			
Tested 12/05/23			
Tester AFRE			
Recorded 12/05/23			
<b>Actual Normality</b>	2.0000	1.9995 - 2.0005	2.0001
	EQ/L		
Sample ID			
Tested 12/05/23			
Tester AFRE			
Recorded 12/05/23			
<b>N.I.S.T. Traceable to SRM 84 (KHP)</b>	YES		YES
	YES		
Sample ID			
Tested 12/05/23			
Tester AFRE			
Recorded 12/05/23			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE LIQUID APPEARANCE	PASS
Sample ID		
Tested	12/05/23	
Tester	AFRE	
Recorded	12/05/23	
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT	M112,VF5-L14
Sample ID		
Tested	12/05/23	
Tester	AFRE	
Recorded	12/05/23	
<b>Lot #</b>	3120140	
<b>Made</b>	12/05/23	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**