



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 3540			
EDTA TITRANT 0.1 MOLAR (0.2N)			
Test	Target/UOM	Range	Result
Actual Molarity	0.1000 MOL/L	0.0995 - 0.1005	0.1002
Sample ID			
Tested	12/12/23		
Tester	MHUS		
Recorded	12/12/23		
QCSP4:STANDARDIZATION EDTA SOLUTIONS	YES QCSP4		YES
Sample ID			
Tested	12/12/23		
Tester	MHUS		
Recorded	12/12/23		
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	12/12/23		
Tester	MHUS		
Recorded	12/12/23		
FORMULA WEIGHT	372.24 F.W.	= 372.24	372.24
Sample ID			
Tested	12/12/23		
Tester	MHUS		
Recorded	12/12/23		
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested	12/12/23		
Tester	MHUS		
Recorded	12/12/23		
Actual Density	1.01671 GM/ML	1.01605 - 1.01736	1.01633
Sample ID			
Tested	12/12/23		
Tester	MHUS		
Recorded	12/12/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M113,T211
Sample ID			
Tested	12/12/23		
Tester	MHUS		
Recorded	12/12/23		
Traceable to N.I.S.T. SRM 915	YES YES		YES
Sample ID			
Tested	12/12/23		
Tester	MHUS		
Recorded	12/12/23		



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

EXPIRATION DATE	MM/DD/YY	12/30/25
	MM/DD/YYYY	
Sample ID		
Tested	12/12/23	
Tester	MHUS	
Recorded	12/12/23	
SHELF LIFE OF PRODUCT	2 YEARS	2 YEARS
	TIME	
Sample ID		
Tested	12/12/23	
Tester	MHUS	
Recorded	12/12/23	
Lot #	3120504	
Made	12/12/23	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager