



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Test	Target/UOM	Range	Result
Packaged Product: SPE509F			
	TOTAL SUSPENDED SOLIDS STANDARD 1000 PPM		
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 12/12/23			
Tester AROS			
Recorded 12/12/23			
APPEARANCE AND COLOR	TURBID LIQUID APPEARANCE		PASS
Sample ID			
Tested 12/12/23			
Tester AROS			
Recorded 12/12/23			
CALCULATED MG/L (PPM) TSS	1000.0 MG/L	995.0 - 1005.0	1000.4
Sample ID			
Tested 12/12/23			
Tester AROS			
Recorded 12/12/23			
ACT GMS CELITE	1.0000 GM	0.9990 - 1.0010	1.0004
Sample ID			
Tested 12/12/23			
Tester AROS			
Recorded 12/12/23			
QCSP11:STANDARDIZATION OF BROMIDE-BROMAT	YES QCSP11		YES
Sample ID			
Tested 12/12/23			
Tester AROS			
Recorded 12/12/23			
MLS OF WATER	1000 ML	995 - 1005	1000
Sample ID			
Tested 12/12/23			
Tester AROS			
Recorded 12/12/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M108,VF1-Z26
Sample ID			
Tested 12/12/23			
Tester AROS			
Recorded 12/12/23			
WEIGHT NIST TRACEABLE	PASS YES		PASS
Sample ID			
Tested 12/12/23			
Tester AROS			
Recorded 12/12/23			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

SHELF LIFE OF PRODUCT	SHELF LIFE TIME	1 YEAR
Sample ID		
Tested 12/12/23		
Tester AROS		
Recorded 12/12/23		
EXPIRATION DATE	MM/DD/YYYY	12/30/2024
	MM/DD/YYYY	
Sample ID		
Tested 12/12/23		
Tester AROS		
Recorded 12/12/23		
Lot # 3121145		
Made 12/12/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager