



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: SGS065F			
250 mg/L 12 Component Mixed Metal Working Solution			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 12/27/23			
Tester MMAT			
Recorded 12/27/23			
N.I.S.T. Traceable	YES		YES
Sample ID	YES		
Tested 12/27/23			
Tester MMAT			
Recorded 12/27/23			
CALCULATED ALUMINUM MG/L NIST	250.00 MG/L	245.00 - 255.00	250.00
Sample ID			
Tested 12/27/23			
Tester MMAT			
Recorded 12/27/23			
CALCULATED SILICON MG/L NIST	250.00 MG/L	245.00 - 255.00	250.00
Sample ID			
Tested 12/27/23			
Tester MMAT			
Recorded 12/27/23			
CALCULATED VANADIUM MG/L NIST	250.00 MG/L	245.00 - 255.00	250.00
Sample ID			
Tested 12/27/23			
Tester MMAT			
Recorded 12/27/23			
CALCULATED NICKEL MG/L NIST	250.00 MG/L	245.00 - 255.00	250.00
Sample ID			
Tested 12/27/23			
Tester MMAT			
Recorded 12/27/23			
CALCULATED IRON MG/L NIST	250.00 MG/L	245.00 - 255.00	250.00
Sample ID			
Tested 12/27/23			
Tester MMAT			
Recorded 12/27/23			
CALCULATED ALUMINUM MG/L NIST	250.00 MG/L	245.00 - 255.00	250.00
Sample ID			
Tested 12/27/23			
Tester MMAT			
Recorded 12/27/23			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

CALCULATED CALCIUM MG/L NIST	250.00	245.00 - 255.00	250.00
	MG/L		
Sample ID			
Tested	12/27/23		
Tester	MMAT		
Recorded	12/27/23		
CALCULATED PHOSPHORUS MG/L NIST	250.00	245.00 - 255.00	250.00
	MG/L		
Sample ID			
Tested	12/27/23		
Tester	MMAT		
Recorded	12/27/23		
CALCULATED ALUMINUM MG/L NIST	250.00	245.00 - 255.00	250.00
	MG/L		
Sample ID			
Tested	12/27/23		
Tester	MMAT		
Recorded	12/27/23		
CALCULATED POTASSIUM MG/L NIST	250.00	245.00 - 255.00	250.00
	MG/L		
Sample ID			
Tested	12/27/23		
Tester	MMAT		
Recorded	12/27/23		
CALCULATED LEAD MG/L NIST	250.00	245.00 - 255.00	250.00
	MG/L		
Sample ID			
Tested	12/27/23		
Tester	MMAT		
Recorded	12/27/23		
CALCULATED MAGNESIUM MG/L NIST	250.00	245.00 - 255.00	250.00
	MG/L		
Sample ID			
Tested	12/27/23		
Tester	MMAT		
Recorded	12/27/23		
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested	12/27/23		
Tester	MMAT		
Recorded	12/27/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P171,P911,VF500-X30
Sample ID			
Tested	12/27/23		
Tester	MMAT		
Recorded	12/27/23		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 12/27/23 at 2:02 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

EXPIRATION DATE

MM/DD/YYYY

12/30/2024

MM/DD/YYYY

Sample ID

Tested 12/27/23

Tester MMAT

Recorded 12/27/23

Lot # 3122062

Made 12/27/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager