



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8375			
SILVER NITRATE 0.01 NORMAL IN IPA (ASTM D3227-16)			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 01/03/24			
Tester SHAR			
Recorded 01/03/24			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 01/03/24			
Tester SHAR			
Recorded 01/03/24			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		01/03/2024
Sample ID			
Tested 01/03/24			
Tester SHAR			
Recorded 01/03/24			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0	10.0 - 35.0	21.0
Sample ID			
Tested 01/03/24			
Tester SHAR			
Recorded 01/03/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		01/03/2024
Sample ID			
Tested 01/03/24			
Tester SHAR			
Recorded 01/03/24			
AMBIENT TEST TEMPERATURE DEGC	20.0	15.0 - 30.0	21.0
Sample ID			
Tested 01/03/24			
Tester SHAR			
Recorded 01/03/24			
Actual Normality	0.0100	0.0097 - 0.0103	0.0097
	EQ/L		
Sample ID			
Tested 01/03/24			
Tester SHAR			
Recorded 01/03/24			
Actual Density	0.79067	0.79000 - 0.79135	0.79054
	GM/ML		
Sample ID			
Tested 01/03/24			
Tester SHAR			
Recorded 01/03/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

QCSP54:STANDARDIZATION OF LOW NORMAL SN	YES QCSP54	YES
Sample ID		
Tested 01/03/24		
Tester SHAR		
Recorded 01/03/24		
N.I.S.T. Traceable to SRM 919 NaCl	YES YES	YES
Sample ID		
Tested 01/03/24		
Tester SHAR		
Recorded 01/03/24		
IPA TESTED FOR PEROXIDES	PASSES YES	PASS
Sample ID		
Tested 01/03/24		
Tester SHAR		
Recorded 01/03/24		
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE	PASS
Sample ID		
Tested 01/03/24		
Tester SHAR		
Recorded 01/03/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M112,VF1-Z39,T13,M55,M81,TI D41124
Sample ID		
Tested 01/03/24		
Tester SHAR		
Recorded 01/03/24		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY	01/30/25
Sample ID		
Tested 01/03/24		
Tester SHAR		
Recorded 01/03/24		
SHELF LIFE OF PRODUCT	SHELF LIFE TIME	1 YEAR
Sample ID		
Tested 01/03/24		
Tester SHAR		
Recorded 01/03/24		

Lot # 3122152

Made 01/03/24



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 01/03/24 at 1:30 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager