



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: UC058F			
SULFURIC ACID 0.5 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 01/20/23			
Tester SNEL			
Recorded 01/20/23			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YERS
Sample ID			
Tested 01/20/23			
Tester SNEL			
Recorded 01/20/23			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		01/20/23
Sample ID			
Tested 01/20/23			
Tester SNEL			
Recorded 01/20/23			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0	10.0 - 35.0	20.4
Sample ID			
Tested 01/20/23			
Tester SNEL			
Recorded 01/20/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		01/20/23
Sample ID			
Tested 01/20/23			
Tester SNEL			
Recorded 01/20/23			
AMBIENT TEST TEMPERATURE DEGC	20.0	15.0 - 30.0	20.6
Sample ID			
Tested 01/20/23			
Tester SNEL			
Recorded 01/20/23			
Actual Normality	0.5000	0.4995 - 0.5005	0.5003
	EQ/L		
Sample ID			
Tested 01/20/23			
Tester SNEL			
Recorded 01/20/23			
QCSP51:STANDARDIZATION HYDROCHLORIC&SULF	YES		YES
Sample ID			
Tested 01/20/23			
Tester SNEL			
Recorded 01/20/23			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

N.I.S.T. Traceable to SRM 84

YES

YES

YES

Sample ID

Tested 01/20/23

Tester SNEL

Recorded 01/20/23

APPEARANCE AND COLOR

CLEAR LIQUID

PASS

APPEARANCE

Sample ID

Tested 01/20/23

Tester SNEL

Recorded 01/20/23

INSTRUMENTS USED DURING

INSTRUMENT

GC,VF5-E5

PREPARATION

Sample ID

Tested 01/20/23

Tester SNEL

Recorded 01/20/23

Lot # 3011833

Made 01/20/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager