



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> LY511F			
Additive Extraction Solution			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 01/24/24			
Tester RSHE			
Recorded 01/24/24			
<b>CALCULATED PPM DIOCTYL PHTHALATE</b>	175.00	170.00 - 180.00	174.88
Sample ID			
Tested 01/24/24			
Tester RSHE			
Recorded 01/24/24			
<b>GRAMS DIOCTYL PHTHALATE</b>	0.5180	0.5100 - 0.5220	0.5192
Sample ID			
Tested 01/24/24			
Tester RSHE			
Recorded 01/24/24			
<b>ACTUAL ASSAY DIOCTYL PHTHALATE</b>	100.0	99.4 - 100.0	99.7
Sample ID			
Tested 01/24/24			
Tester RSHE			
Recorded 01/24/24			
<b>TOTAL COMBINED WT IN GMS</b>	2960.0	2940.0 - 2980.0	2960.0
Sample ID			
Tested 01/24/24			
Tester RSHE			
Recorded 01/24/24			
<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 01/24/24			
Tester RSHE			
Recorded 01/24/24			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		VF2-A9,GC,M115,M03
Sample ID			
Tested 01/24/24			
Tester RSHE			
Recorded 01/24/24			
<b>EXPIRATION DATE</b>	MM/DD/YYYY		01/30/2025
Sample ID			
Tested 01/24/24			
Tester RSHE			
Recorded 01/24/24			



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 2

Date: 01/24/24 at 7:04 AM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

**Lot #** 40122132

**Made** 01/24/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**