



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8754			
SODIUM STANDARD 1.0 PPM N.I.S.T. TRACEABLE SOLUTION			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 01/11/24			
Tester AROS			
Recorded 01/11/24			
PPM SODIUM	1.00	0.98 - 1.02	1.00
	PPM		
Sample ID			
Tested 01/11/24			
Tester AROS			
Recorded 01/11/24			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 01/11/24			
Tester AROS			
Recorded 01/11/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P135,VF1-Z39
Sample ID			
Tested 01/11/24			
Tester AROS			
Recorded 01/11/24			
SHELF LIFE OF PRODUCT	6 MONTHS TIME		6 MONTHS
Sample ID			
Tested 01/11/24			
Tester AROS			
Recorded 01/11/24			
EXPIRATION DATE	MM/DD/YYYY		07/30/2024
	MM/DD/YYYY		
Sample ID			
Tested 01/11/24			
Tester AROS			
Recorded 01/11/24			
Lot #	4011132		
Made	01/11/24		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 01/11/24 at 5:17 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Quality Manager