



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| Packaged Product: 4130 | | | |
|--|------------|-----------------|------------|
| HYDROCHLORIC ACID 0.1 NORMAL | | | |
| Test | Target/UOM | Range | Result |
| PREPARED TO FORMULATION ON FILE | YES | | YES |
| Sample ID | | | |
| Tested 01/30/24 | | | |
| Tester DTRA | | | |
| Recorded 01/30/24 | | | |
| SAMPLE COLLECTION & HANDLING PER LAB040 | YES | | YES |
| Sample ID | | | |
| Tested 01/30/24 | | | |
| Tester DTRA | | | |
| Recorded 01/30/24 | | | |
| DATE OF SAMPLING FOR TEST PURPOSES | MM/DD/YYYY | | 01/30/2024 |
| Sample ID | MM/DD/YYYY | | |
| Tested 01/30/24 | | | |
| Tester DTRA | | | |
| Recorded 01/30/24 | | | |
| AMBIENT TEMPERATURE DURING SAMPLING DEGC | 25.0 | 10.0 - 35.0 | 21.0 |
| Sample ID | DEG C | | |
| Tested 01/30/24 | | | |
| Tester DTRA | | | |
| Recorded 01/30/24 | | | |
| DATE OF TESTING OF SAMPLE | MM/DD/YYYY | | 01/30/2024 |
| Sample ID | MM/DD/YYYY | | |
| Tested 01/30/24 | | | |
| Tester DTRA | | | |
| Recorded 01/30/24 | | | |
| AMBIENT TEST TEMPERATURE DEGC | 20.0 | 15.0 - 30.0 | 21.0 |
| Sample ID | DEG C | | |
| Tested 01/30/24 | | | |
| Tester DTRA | | | |
| Recorded 01/30/24 | | | |
| Actual Normality | 0.1000 | 0.0995 - 0.1005 | 0.0996 |
| Sample ID | EQ/L | | |
| Tested 01/30/24 | | | |
| Tester DTRA | | | |
| Recorded 01/30/24 | | | |
| + Measurement Uncertainty | 0.0008 | = 0.0008 | 0.0008 |
| Sample ID | EQ/L | | |
| Tested 01/30/24 | | | |
| Tester DTRA | | | |
| Recorded 01/30/24 | | | |



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| | | | |
|--|----------------------------|-------------------|------------|
| QCSP5: STANDARDIZATION OF AQUEOUS ACIDS | YES METHOD | | YES |
| Sample ID | | | |
| Tested | 01/30/24 | | |
| Tester | DTRA | | |
| Recorded | 01/30/24 | | |
| N.I.S.T. Traceable to SRM 723 | YES YES | | YES |
| Sample ID | | | |
| Tested | 01/30/24 | | |
| Tester | DTRA | | |
| Recorded | 01/30/24 | | |
| Actual Density | 1.00008 GM/ML | 0.99995 - 1.00021 | 1.00009 |
| Sample ID | | | |
| Tested | 01/30/24 | | |
| Tester | DTRA | | |
| Recorded | 01/30/24 | | |
| APPEARANCE AND COLOR | CLEAR LIQUID APPEARANCE | | PASS |
| Sample ID | | | |
| Tested | 01/30/24 | | |
| Tester | DTRA | | |
| Recorded | 01/30/24 | | |
| INSTRUMENTS USED DURING PREPARATION | INSTRUMENT | | M113,T212 |
| Sample ID | | | |
| Tested | 01/30/24 | | |
| Tester | DTRA | | |
| Recorded | 01/30/24 | | |
| EXPIRATION DATE | MM/DD/YYYY | | 02/28/2026 |
| Sample ID | | | |
| Tested | 01/30/24 | | |
| Tester | DTRA | | |
| Recorded | 01/30/24 | | |
| SHELF LIFE OF PRODUCT | 2 YEARS TIME | | 2 YEARS |
| Sample ID | | | |
| Tested | 01/30/24 | | |
| Tester | DTRA | | |
| Recorded | 01/30/24 | | |

Lot # 4012555

Made 01/30/24



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 01/30/24 at 11:52 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager