



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 6487			
NITRITE STANDARD 300 PPM (NO2)			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 01/31/24			
Tester MMAT			
Recorded 01/31/24			
PPM NITRITE AS NO2	300.0	296.0 - 304.0	300.0
	PPM		
Sample ID			
Tested 01/31/24			
Tester MMAT			
Recorded 01/31/24			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 01/31/24			
Tester MMAT			
Recorded 01/31/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M115,VF2-B14
Sample ID			
Tested 01/31/24			
Tester MMAT			
Recorded 01/31/24			
EXPIRATION DATE	MM/DD/YYYY		05/30/2024
	MM/DD/YYYY		
Sample ID			
Tested 01/31/24			
Tester MMAT			
Recorded 01/31/24			
Lot #	4013121		
Made	01/31/24		

Quality Manager