



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| Packaged Product: HAC460F | | | |
|--|------------------------------|-------------|-------------|
| Calibration Soln 0.5 PPM CaCO3 for EZ1016-Total Hardness | | | |
| Test | Target/UOM | Range | Result |
| PREPARED TO FORMULATION ON | YES | | YES |
| FILE | YES | | |
| Sample ID | | | |
| Tested 01/29/24 | | | |
| Tester MHAY | | | |
| Recorded 01/29/24 | | | |
| CALCULATED PPM CaCO3 | 0.50 | 0.47 - 0.53 | 0.50 |
| | PPM | | |
| Sample ID | | | |
| Tested 01/29/24 | | | |
| Tester MHAY | | | |
| Recorded 01/29/24 | | | |
| APPEARANCE AND COLOR | CLEAR WATER WHITE APPEARANCE | | PASS |
| Sample ID | | | |
| Tested 01/29/24 | | | |
| Tester MHAY | | | |
| Recorded 01/29/24 | | | |
| INSTRUMENTS USED DURING PREPARATION | INSTRUMENT | | P172,VF2-B3 |
| Sample ID | | | |
| Tested 01/30/24 | | | |
| Tester SHAR | | | |
| Recorded 01/30/24 | | | |
| EXPIRATION DATE | MM/DD/YYYY | | 05/30/2024 |
| | MM/DD/YYYY | | |
| Sample ID | | | |
| Tested 01/30/24 | | | |
| Tester SHAR | | | |
| Recorded 01/30/24 | | | |
| Lot # | 4012659 | | |
| Made | 01/30/24 | | |

Richard [Signature]
Quality Manager