



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8635			
SODIUM HYDROXIDE 0.25 NORMAL NIST TRACEABLE SOLUTION			
Test	Target/UOM	Range	Result
Actual Normality	0.2500 EQ/L	0.2495 - 0.2505	0.2504
Sample ID			
Tested	01/18/24		
Tester	ENOR		
Recorded	01/18/24		
QCSP7: STANDARDIZATION OF AQUEOUS BASES	YES METHOD		YES
Sample ID			
Tested	01/18/24		
Tester	ENOR		
Recorded	01/18/24		
N.I.S.T. Traceable to SRM 919	Yes YES		YES
Sample ID			
Tested	01/18/24		
Tester	ENOR		
Recorded	01/18/24		
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	01/18/24		
Tester	ENOR		
Recorded	01/18/24		
SAMPLE COLLECTION & HANDLING PER LAB040	YES YES		YES
Sample ID			
Tested	01/18/24		
Tester	ENOR		
Recorded	01/18/24		
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY MM/DD/YYYY		01/18/2024
Sample ID			
Tested	01/18/24		
Tester	ENOR		
Recorded	01/18/24		
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	20.3
Sample ID			
Tested	01/18/24		
Tester	ENOR		
Recorded	01/18/24		
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		01/18/2024
Sample ID			
Tested	01/18/24		
Tester	ENOR		
Recorded	01/18/24		



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	20.0
DEG C	DEG C		
Sample ID			
Tested	01/18/24		
Tester	ENOR		
Recorded	01/18/24		
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested	01/18/24		
Tester	ENOR		
Recorded	01/18/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M103,T-169,M55,BID08251,PID M91
Sample ID			
Tested	01/18/24		
Tester	ENOR		
Recorded	01/18/24		
EXPIRATION DATE	MM/DD/YY		01/30/26
	MM/DD/YYYY		
Sample ID			
Tested	01/18/24		
Tester	ENOR		
Recorded	01/18/24		
SHELF LIFE OF PRODUCT	2 YEARS		2 YEARS
	TIME		
Sample ID			
Tested	01/18/24		
Tester	ENOR		
Recorded	01/18/24		
Actual Density	1.00939	1.00925 - 1.00953	1.00951
	GM/ML		
Sample ID			
Tested	01/18/24		
Tester	ENOR		
Recorded	01/18/24		
Lot #	4011712		
Made	01/18/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 01/18/24 at 11:17 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager