



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 3950			
GLUCOSE-GLUTAMIC ACID SOLUTION			
Test	Target/UOM	Range	Result
ACTUAL PPM CERTIFIED TOC VALUE	122 PPM	112 - 132	121
Sample ID Tested 01/18/24 Tester MHAY Recorded 01/18/24			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID Tested 01/18/24 Tester MHAY Recorded 01/18/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M115,VF1-Z43
Sample ID Tested 01/18/24 Tester MHAY Recorded 01/18/24			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		02/08/24
Sample ID Tested 01/18/24 Tester MHAY Recorded 01/18/24			
SHELF LIFE	3WKS TIME		3 WEEKS
Sample ID Tested 01/18/24 Tester MHAY Recorded 01/18/24			
Lot #	4011860		
Made	01/18/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.