



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 4086			
HYDROCHLORIC ACID 0.005 NORMAL IN METHANOL			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 01/24/24			
Tester SHAR			
Recorded 01/24/24			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID			
Tested 01/24/24			
Tester SHAR			
Recorded 01/24/24			
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY		01/24/2024
Sample ID			
Tested 01/24/24			
Tester SHAR			
Recorded 01/24/24			
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0	10.0 - 35.0	21.0
Sample ID			
Tested 01/24/24			
Tester SHAR			
Recorded 01/24/24			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY		01/24/2024
Sample ID			
Tested 01/24/24			
Tester SHAR			
Recorded 01/24/24			
<b>AMBIENT TEST TEMPERATURE DEGC</b>	20.0	15.0 - 30.0	21.0
Sample ID			
Tested 01/24/24			
Tester SHAR			
Recorded 01/24/24			
<b>NORMALITY TO 5 DECIMAL PLACES</b>	0.00500	0.00470 - 0.00530	0.00476
Sample ID			
Tested 01/24/24			
Tester SHAR			
Recorded 01/24/24			
<b>QCSP5: STANDARDIZATION OF AQUEOUS ACIDS</b>	YES		YES
Sample ID			
Tested 01/24/24			
Tester SHAR			
Recorded 01/24/24			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>Actual Density</b>	0.79195 GM/ML	0.79120 - 0.79269	0.79195
Sample ID			
Tested	01/24/24		
Tester	SHAR		
Recorded	01/24/24		
<b>N.I.S.T. Traceable to SRM 723 THAM</b>	YES YES		YES
Sample ID			
Tested	01/24/24		
Tester	SHAR		
Recorded	01/24/24		
<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested	01/24/24		
Tester	SHAR		
Recorded	01/24/24		
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M55,M112,EF2-SD
Sample ID			
Tested	01/24/24		
Tester	SHAR		
Recorded	01/24/24		
<b>SHELF LIFE OF PRODUCT</b>	4 MONTHS TIME		4 MONTHS
Sample ID			
Tested	01/24/24		
Tester	SHAR		
Recorded	01/24/24		
<b>EXPIRATION DATE</b>	MM/DD/YYYY MM/DD/YYYY		05/30/2024
Sample ID			
Tested	01/24/24		
Tester	SHAR		
Recorded	01/24/24		
<b>Lot #</b>	4012428		
<b>Made</b>	01/24/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 3

Date: 01/24/24 at 2:34 PM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**