



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| Packaged Product: PR017F | | | |
|--|-------------|-----------------|------------|
| Hydrochloric acid 0.1 Normal Pheno-EP | | | |
| Test | Target/UOM | Range | Result |
| PREPARED TO FORMULATION ON FILE | YES | | YES |
| Sample ID | YES | | |
| Tested 02/01/24 | | | |
| Tester JENC | | | |
| Recorded 02/01/24 | | | |
| SAMPLE COLLECTION & HANDLING PER LAB040 | YES | | YES |
| Sample ID | YES | | |
| Tested 02/01/24 | | | |
| Tester JENC | | | |
| Recorded 02/01/24 | | | |
| DATE OF SAMPLING FOR TEST PURPOSES | MM/DD/YYYY | | 02/01/2024 |
| Sample ID | MM/DD/YYYY | | |
| Tested 02/01/24 | | | |
| Tester JENC | | | |
| Recorded 02/01/24 | | | |
| AMBIENT TEMPERATURE DURING SAMPLING DEGC | 25.0 | 10.0 - 35.0 | 21.5 |
| Sample ID | DEG C | | |
| Tested 02/01/24 | | | |
| Tester JENC | | | |
| Recorded 02/01/24 | | | |
| DATE OF TESTING OF SAMPLE | MM/DD/YYYY | | 02/01/2024 |
| Sample ID | MM/DD/YYYY | | |
| Tested 02/01/24 | | | |
| Tester JENC | | | |
| Recorded 02/01/24 | | | |
| AMBIENT TEST TEMPERATURE DEG C | 20.0 | 15.0 - 30.0 | 21.5 |
| Sample ID | DEG C | | |
| Tested 02/01/24 | | | |
| Tester JENC | | | |
| Recorded 02/01/24 | | | |
| NORMALITY TO 4 DEC PLACES | 0.1000 | 0.0995 - 0.1005 | 0.1004 |
| Sample ID | EQUIV/LITER | | |
| Tested 02/01/24 | | | |
| Tester JENC | | | |
| Recorded 02/01/24 | | | |
| STANDARDIZED TO A PHENOLPHTHALEIN EP | YES | | YES |
| Sample ID | YES | | |
| Tested 02/01/24 | | | |
| Tester JENC | | | |
| Recorded 02/01/24 | | | |



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

APPEARANCE AND COLOR

CLEAR LIQUID
 APPEARANCE

PASS

Sample ID
 Tested 02/01/24
 Tester JENC
 Recorded 02/01/24

**INSTRUMENTS USED DURING
 PREPARATION**

INSTRUMENT

GC,T-227,BID06419,PID-121

Sample ID
 Tested 02/01/24
 Tester JENC
 Recorded 02/01/24

Lot # 4012656

Made 02/01/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager