



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: SPX475F			
NITROGEN STANDARD 1000 PPM AS N AQUEOUS SOLUTION			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 01/15/24			
Tester AROS			
Recorded 01/15/24			
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested 01/15/24			
Tester AROS			
Recorded 01/15/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M108,VF1-Z24
Sample ID			
Tested 01/15/24			
Tester AROS			
Recorded 01/15/24			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		03/20/24
Sample ID			
Tested 01/15/24			
Tester AROS			
Recorded 01/15/24			
PPM NITROGEN NO DECIAML PLACES	1000 PPM	985 - 1015	999
Sample ID			
Tested 01/15/24			
Tester AROS			
Recorded 01/15/24			
Lot #	4011280		
Made	01/15/24		

Quality Manager