



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Test	Target/UOM	Range	Result
Packaged Product: 4191			
	HYDROCHLORIC ACID 0.5 NORMAL IN IPA		
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 02/06/24			
Tester JENC			
Recorded 02/06/24			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 02/06/24			
Tester JENC			
Recorded 02/06/24			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		02/06/2024
Sample ID			
Tested 02/06/24			
Tester JENC			
Recorded 02/06/24			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	21.7
Sample ID			
Tested 02/06/24			
Tester JENC			
Recorded 02/06/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		02/06/2024
Sample ID			
Tested 02/06/24			
Tester JENC			
Recorded 02/06/24			
AMBIENT TEST TEMPERATURE DEGC	20.0 DEG C	15.0 - 30.0	21.7
Sample ID			
Tested 02/06/24			
Tester JENC			
Recorded 02/06/24			
N.I.S.T. Traceable to SRM 723	YES YES		YES
Sample ID			
Tested 02/06/24			
Tester JENC			
Recorded 02/06/24			
QCSP5: STANDARDIZATION OF AQUEOUS ACIDS	YES METHOD		YES
Sample ID			
Tested 02/06/24			
Tester JENC			
Recorded 02/06/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

Actual Normality	0.5000 EQ/L	0.4995 - 0.5005	0.5000
Sample ID			
Tested	02/06/24		
Tester	JENC		
Recorded	02/06/24		
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested	02/06/24		
Tester	JENC		
Recorded	02/06/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M87,T-228,M55,BURET:02259, M91
Sample ID			
Tested	02/06/24		
Tester	JENC		
Recorded	02/06/24		
Actual Density	0.81013 GM/ML	0.80934 - 0.81092	0.80998
Sample ID			
Tested	02/06/24		
Tester	JENC		
Recorded	02/06/24		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		02/28/25
Sample ID			
Tested	02/06/24		
Tester	JENC		
Recorded	02/06/24		
Lot #	4020202		
Made	02/06/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager