



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: LY396F			
ORGANIC ACID CALIBRATION STOCK 1,000 PPM EA			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 02/06/24			
Tester GWEB			
Recorded 02/06/24			
CALCULATED PPM PROPIONATE	1000.00	995.00 - 1005.00	999.99
	PPM		
Sample ID			
Tested 02/06/24			
Tester GWEB			
Recorded 02/06/24			
CALCULATED PPM ACETATE	1000.00	995.00 - 1005.00	997.02
	PPM		
Sample ID			
Tested 02/06/24			
Tester GWEB			
Recorded 02/06/24			
CALCULATED PPM FORMATE	1000.00	995.00 - 1005.00	1004.01
	PPM		
Sample ID			
Tested 02/06/24			
Tester GWEB			
Recorded 02/06/24			
CALCULATED PPM BUTYRATE	1000.00	995.00 - 1005.00	1000.14
	PPM		
Sample ID			
Tested 02/06/24			
Tester GWEB			
Recorded 02/06/24			
CALCULATED PPM CARBONATE	1000.00	995.00 - 1005.00	998.49
	PPM		
Sample ID			
Tested 02/06/24			
Tester GWEB			
Recorded 02/06/24			
CALCULATED PPM NITRATE	1000.00	995.00 - 1005.00	999.64
	PPM		
Sample ID			
Tested 02/06/24			
Tester GWEB			
Recorded 02/06/24			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 02/06/24			
Tester GWEB			
Recorded 02/06/24			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 02/06/24 at 8:10 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M108,VF1-Z16
Sample ID Tested 02/06/24 Tester GWEB Recorded 02/06/24		
Lot # 4020584 Made 02/06/24		

Quality Manager