



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

Test	Target/UOM	Range	Result
<b>Packaged Product:</b> SPE512F			
	POTASSIUM HYDROXIDE 5.0 NORMAL (STANDARDIZED)		
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 02/09/24			
Tester SHAR			
Recorded 02/09/24			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID			
Tested 02/09/24			
Tester SHAR			
Recorded 02/09/24			
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY		02/09/2024
Sample ID			
Tested 02/09/24			
Tester SHAR			
Recorded 02/09/24			
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0 DEG C	10.0 - 35.0	21.0
Sample ID			
Tested 02/09/24			
Tester SHAR			
Recorded 02/09/24			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY MM/DD/YYYY		02/09/2024
Sample ID			
Tested 02/09/24			
Tester SHAR			
Recorded 02/09/24			
<b>AMBIENT TEST TEMPERATURE DEGC</b>	20.0 DEG C	15.0 - 30.0	21.0
Sample ID			
Tested 02/09/24			
Tester SHAR			
Recorded 02/09/24			
<b>NORMALITY TO 4 DEC PLACES</b>	5.0000 EQUIV/LITER	4.9800 - 5.0200	4.9930
Sample ID			
Tested 02/09/24			
Tester SHAR			
Recorded 02/09/24			
<b>N.I.S.T. Traceable to SRM 84</b>	YES		YES
Sample ID			
Tested 02/09/24			
Tester SHAR			
Recorded 02/09/24			



Aqua Solutions, Inc.  
 Ph: 281-479-2569  
 Fax: 281-479-2790  
 P.O. Box 70  
 Deer Park, TX 77536  
 Website: [www.aquasolutions.org](http://www.aquasolutions.org)

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID APPEARANCE	PASS
Sample ID		
Tested	02/09/24	
Tester	SHAR	
Recorded	02/09/24	
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT	M87,T-220,BID-079,PID-M91
Sample ID		
Tested	02/09/24	
Tester	SHAR	
Recorded	02/09/24	
<b>Lot #</b>	4020855	
<b>Made</b>	02/09/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**