



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 4095			
HYDROCHLORIC ACID 0.01 NORMAL IN METHANOL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 02/19/24			
Tester MHUS			
Recorded 02/19/24			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID	YES		
Tested 02/19/24			
Tester MHUS			
Recorded 02/19/24			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		02/19/2024
Sample ID	MM/DD/YYYY		
Tested 02/19/24			
Tester MHUS			
Recorded 02/19/24			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0	10.0 - 35.0	22.0
Sample ID	DEG C		
Tested 02/19/24			
Tester MHUS			
Recorded 02/19/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		02/19/2024
Sample ID	MM/DD/YYYY		
Tested 02/19/24			
Tester MHUS			
Recorded 02/19/24			
AMBIENT TEST TEMPERATURE DEG C	20.0	15.0 - 30.0	22.0
Sample ID	DEG C		
Tested 02/19/24			
Tester MHUS			
Recorded 02/19/24			
Actual Normality	0.0100	0.0095 - 0.0105	0.0098
Sample ID	EQ/L		
Tested 02/19/24			
Tester MHUS			
Recorded 02/19/24			
QCSP5: STANDARDIZATION OF AQUEOUS ACIDS	YES		YES
Sample ID	METHOD		
Tested 02/19/24			
Tester MHUS			
Recorded 02/19/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

N.I.S.T. Traceable to SRM 723 THAM	YES	YES
	YES	
Sample ID		
Tested	02/19/24	
Tester	MHUS	
Recorded	02/19/24	
Actual Density	0.79222	0.79129 - 0.79315
	GM/ML	
Sample ID		
Tested	02/19/24	
Tester	MHUS	
Recorded	02/19/24	
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE	PASS
Sample ID		
Tested	02/19/24	
Tester	MHUS	
Recorded	02/19/24	
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	GC,T-44,BID-04099,PID-M91
Sample ID		
Tested	02/19/24	
Tester	MHUS	
Recorded	02/19/24	
EXPIRATION DATE	MM/DD/YYYY	08/30/2024
	MM/DD/YYYY	
Sample ID		
Tested	02/19/24	
Tester	MHUS	
Recorded	02/19/24	
Lot #	4021665	
Made	02/19/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager