



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Test	Target/UOM	Range	Result
Packaged Product: OXY016F			
	FERROZINE COLOR REAGENT * YELLOW SOLUTION *		
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 02/23/24			
Tester BPHI			
Recorded 02/23/24			
GPL SODIUM ACETATE TRIHYDRATE	550.0	549.0 - 551.0	550.0
Sample ID	GPL		
Tested 02/23/24			
Tester BPHI			
Recorded 02/23/24			
MLS/LITER ACETIC ACID	50.0	49.9 - 50.5	50.0
Sample ID	ML/L		
Tested 02/23/24			
Tester BPHI			
Recorded 02/23/24			
CALCULATED GPL FERROZINE	1.250	1.245 - 1.255	1.250
Sample ID	GPL		
Tested 02/23/24			
Tester BPHI			
Recorded 02/23/24			
Actual Density	1.16077	1.15711 - 1.16443	1.16272
Sample ID	GM/ML		
Tested 02/23/24			
Tester BPHI			
Recorded 02/23/24			
APPEARANCE AND COLOR	CLEAR YELLOW LIQUID APPEARANCE		PASS
Sample ID			
Tested 02/23/24			
Tester BPHI			
Recorded 02/23/24			
INSTRUMENTS USED DURING PREPARATION	LAB EQUIPMENT INSTRUMENT		M87,GC,T-204,M55
Sample ID			
Tested 02/23/24			
Tester BPHI			
Recorded 02/23/24			
EXPIRATION DATE	MM/DD/YYYY		02/28/2025
Sample ID	MM/DD/YYYY		
Tested 02/23/24			
Tester BPHI			
Recorded 02/23/24			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 02/23/24 at 9:49 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Lot #	4022335
Made	02/23/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager