



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 4353			
HYDROCHLORIC ACID 1.1 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	1.1000	1.0995 - 1.1008	1.0996
	EQ/L		
Sample ID			
Tested	02/27/24		
Tester	MHUS		
Recorded	02/27/24		
QCSP5: STANDARDIZATION OF AQUEOUS ACIDS	YES		YES
	METHOD		
Sample ID			
Tested	02/27/24		
Tester	MHUS		
Recorded	02/27/24		
N.I.S.T. Traceable to SRM 84 (KHP)	Yes		YES
	YES		
Sample ID			
Tested	02/27/24		
Tester	MHUS		
Recorded	02/27/24		
PREPARED TO FORMULATION ON FILE	YES		YES
	YES		
Sample ID			
Tested	02/27/24		
Tester	MHUS		
Recorded	02/27/24		
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
	YES		
Sample ID			
Tested	02/27/24		
Tester	MHUS		
Recorded	02/27/24		
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		02/27/2024
	MM/DD/YYYY		
Sample ID			
Tested	02/27/24		
Tester	MHUS		
Recorded	02/27/24		
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0	10.0 - 35.0	22.0
	DEG C		
Sample ID			
Tested	02/27/24		
Tester	MHUS		
Recorded	02/27/24		
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		02/27/2024
	MM/DD/YYYY		
Sample ID			
Tested	02/27/24		
Tester	MHUS		
Recorded	02/27/24		



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	22.0
DEG C	DEG C		
Sample ID			
Tested	02/27/24		
Tester	MHUS		
Recorded	02/27/24		
APPEARANCE AND COLOR	CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested	02/27/24		
Tester	MHUS		
Recorded	02/27/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M112,VF5-E4,M55,BURET:390
			2
Sample ID			
Tested	02/27/24		
Tester	MHUS		
Recorded	02/27/24		
EXPIRATION DATE	MM/DD/YYYY		02/28/2026
	MM/DD/YYYY		
Sample ID			
Tested	02/27/24		
Tester	MHUS		
Recorded	02/27/24		
Actual Density	1.01764	1.01760 - 1.01769	1.01760
	GM/ML		
Sample ID			
Tested	02/27/24		
Tester	MHUS		
Recorded	02/27/24		
Lot #	4022644		
Made	02/27/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager