



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: ERL125F			
SILVER NITRATE 0.1 NORMAL IN ISOPROPANOL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 02/27/24			
Tester MHUS			
Recorded 02/27/24			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 02/27/24			
Tester MHUS			
Recorded 02/27/24			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		02/27/2024
Sample ID			
Tested 02/27/24			
Tester MHUS			
Recorded 02/27/24			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0	10.0 - 35.0	22.0
Sample ID			
Tested 02/27/24			
Tester MHUS			
Recorded 02/27/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		02/27/2024
Sample ID			
Tested 02/27/24			
Tester MHUS			
Recorded 02/27/24			
AMBIENT TEST TEMPERATURE DEGC	20.0	15.0 - 30.0	22.0
Sample ID			
Tested 02/27/24			
Tester MHUS			
Recorded 02/27/24			
Actual Normality	0.1000	0.0995 - 0.1005	0.1007
	EQ/L		
Sample ID			
Tested 02/27/24			
Tester MHUS			
Recorded 02/27/24			
QCSP53: SILVER NITRATE STD ON AUTO-TITRA	YES		YES
	METHOD		
Sample ID			
Tested 02/27/24			
Tester MHUS			
Recorded 02/27/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

N.I.S.T. Traceable to SRM 919 NaCl	YES YES	YES
Sample ID		
Tested	02/27/24	
Tester	MHUS	
Recorded	02/27/24	
PREPARED TO ASTM D3227	YES YES	YES
Sample ID		
Tested	02/27/24	
Tester	MHUS	
Recorded	02/27/24	
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE	PASS
Sample ID		
Tested	02/27/24	
Tester	MHUS	
Recorded	02/27/24	
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	EF1-FM,M55,P81
Sample ID		
Tested	02/27/24	
Tester	MHUS	
Recorded	02/27/24	
LOT NUMBER OF ORIGINAL MATERIAL	LOT #	3103015
Sample ID		
Tested	02/27/24	
Tester	MHUS	
Recorded	02/27/24	
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY	03/12/24
Sample ID		
Tested	02/27/24	
Tester	MHUS	
Recorded	02/27/24	
SHELF LIFE OF PRODUCT	1 WEEK TIME	1 WEEK
Sample ID		
Tested	02/27/24	
Tester	MHUS	
Recorded	02/27/24	

Lot # 4022608

Made 02/27/24



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 02/27/24 at 11:13 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager