



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: HON012F			
Calibration Std #3 100.0 mg/L Cl Solution			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 02/29/24			
Tester MMAT			
Recorded 02/29/24			
Actual ppm Fluoride	100.0	99.5 - 100.5	100.0
	PPM		
Sample ID			
Tested 02/29/24			
Tester MMAT			
Recorded 02/29/24			
PPM CHLORIDE	100.00	99.50 - 100.50	100.00
	PPM		
Sample ID			
Tested 02/29/24			
Tester MMAT			
Recorded 02/29/24			
PPM SULFATE	100.00	99.50 - 100.50	100.00
	PPM		
Sample ID			
Tested 02/29/24			
Tester MMAT			
Recorded 02/29/24			
APPEARANCE AND COLOR	CLEAR		PASS
	APPEARANCE		
Sample ID			
Tested 02/29/24			
Tester MMAT			
Recorded 02/29/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P171,VF1-Z31
Sample ID			
Tested 02/29/24			
Tester MMAT			
Recorded 02/29/24			
Lot #	4022969		
Made	02/29/24		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536

Website: www.aquasolutions.org

Page: 2

Date: 02/29/24 at 4:26 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Quality Manager