



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 7243			
POTASSIUM HYDROXIDE 0.05 NORMAL IN METHANOL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 03/08/24			
Tester JENC			
Recorded 03/08/24			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID	YES		
Tested 03/08/24			
Tester JENC			
Recorded 03/08/24			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		03/08/2024
Sample ID	MM/DD/YYYY		
Tested 03/08/24			
Tester JENC			
Recorded 03/08/24			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0	10.0 - 35.0	21.7
Sample ID	DEG C		
Tested 03/08/24			
Tester JENC			
Recorded 03/08/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		03/08/2024
Sample ID	MM/DD/YYYY		
Tested 03/08/24			
Tester JENC			
Recorded 03/08/24			
AMBIENT TEST TEMPERATURE DEG C	20.0	15.0 - 30.0	21.7
Sample ID	DEG C		
Tested 03/08/24			
Tester JENC			
Recorded 03/08/24			
Actual Normality	0.0500	0.0495 - 0.0505	0.0503
Sample ID	EQ/L		
Tested 03/08/24			
Tester JENC			
Recorded 03/08/24			
QCSP7: STANDARDIZATION OF AQUEOUS BASES	YES		YES
Sample ID	METHOD		
Tested 03/08/24			
Tester JENC			
Recorded 03/08/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

N.I.S.T. Traceable to SRM 84 (KHP)	YES	YES
	YES	
Sample ID		
Tested	03/08/24	
Tester	JENC	
Recorded	03/08/24	
Actual Density	0.79629	0.79559 - 0.79700
	GM/ML	
Sample ID		
Tested	03/08/24	
Tester	JENC	
Recorded	03/08/24	
APPEARANCE AND COLOR	CLEAR WATER WHITE	PASS
	LIQUID	
	APPEARANCE	
Sample ID		
Tested	03/08/24	
Tester	JENC	
Recorded	03/08/24	
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M112,T-157,M55,BID06419,PID M91
Sample ID		
Tested	03/08/24	
Tester	JENC	
Recorded	03/08/24	
EXPIRATION DATE	MM/DD/YYYY	03/30/2025
	MM/DD/YYYY	
Sample ID		
Tested	03/08/24	
Tester	JENC	
Recorded	03/08/24	
Lot #	4030611	
Made	03/08/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager