



HALLIBURTON ENERGY SERV./FANN INST.  
AP DEPT./HALLIBURTON ENGY SERV, P.O. BOX 4960  
Portland, OR, 97208-4960  
2818714000

**Certificate of Analysis / QC Results**

Customer PO # : 4517418387

<b>Packaged Product:</b>	209890 BROMOCRESOL GREEN-METHYL RED INDICATOR SOLUTION (2OZ)		
<b>Lot #</b>	4030644		
<b>Chemical Product:</b>	209890F BROMOCRESOL GREEN-METHYL RED INDICATOR SOLUTION		
<b>Test</b>	<b>Target/UOM</b>	<b>Range</b>	<b>Result</b>
PREPARED TO FORMULATION ON FILE	YES		YES
APPEARANCE AND COLOR	DARK GREEN LIQUID APPEARANCE		PASS
INSTRUMENTS USED DURING PREPARATION	LAB EQUIPMENT INSTRUMENT		M108,GC,ER1-DOG
EXPIRATION DATE	MM/DD/YYYY		03/30/2026
<b>Made</b>	03/07/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

This report shall not be reproduced, except in full, without approval of the laboratory.

Aqua Solutions' Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by DNV Certificate Number 51119-2009-AQ-USA-ANAB.

**Quality Manager**  
Aqua Solutions, Inc.