



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> ERL125F			
SILVER NITRATE 0.1 NORMAL IN ISOPROPANOL			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 03/07/24			
Tester BPHI			
Recorded 03/07/24			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID			
Tested 03/07/24			
Tester BPHI			
Recorded 03/07/24			
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY		03/07/2024
Sample ID			
Tested 03/07/24			
Tester BPHI			
Recorded 03/07/24			
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0	10.0 - 35.0	21.1
Sample ID			
Tested 03/07/24			
Tester BPHI			
Recorded 03/07/24			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY		03/07/2024
Sample ID			
Tested 03/07/24			
Tester BPHI			
Recorded 03/07/24			
<b>AMBIENT TEST TEMPERATURE DEGC</b>	20.0	15.0 - 30.0	21.1
Sample ID			
Tested 03/07/24			
Tester BPHI			
Recorded 03/07/24			
<b>Actual Normality</b>	0.1000	0.0995 - 0.1005	0.1003
	EQ/L		
Sample ID			
Tested 03/07/24			
Tester BPHI			
Recorded 03/07/24			
<b>QCSP53: SILVER NITRATE STD ON AUTO-TITRA</b>	YES		YES
	METHOD		
Sample ID			
Tested 03/07/24			
Tester BPHI			
Recorded 03/07/24			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>N.I.S.T. Traceable to SRM 919 NaCl</b>	YES		YES
	YES		
Sample ID			
Tested	03/07/24		
Tester	BPHI		
Recorded	03/07/24		
<b>PREPARED TO ASTM D3227</b>	YES		YES
	YES		
Sample ID			
Tested	03/07/24		
Tester	BPHI		
Recorded	03/07/24		
<b>Actual Density</b>	0.82606	0.82377 - 0.82832	0.82642
	GM/ML		
Sample ID			
Tested	03/07/24		
Tester	BPHI		
Recorded	03/07/24		
<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested	03/07/24		
Tester	BPHI		
Recorded	03/07/24		
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M55,T18,M103,GC
Sample ID			
Tested	03/07/24		
Tester	BPHI		
Recorded	03/07/24		
<b>EXPIRATION DATE</b>	MM/DD/YY		03/19/24
	MM/DD/YYYY		
Sample ID			
Tested	03/07/24		
Tester	BPHI		
Recorded	03/07/24		
<b>SHELF LIFE OF PRODUCT</b>	1 WEEK		1 WEEK
	TIME		
Sample ID			
Tested	03/07/24		
Tester	BPHI		
Recorded	03/07/24		

**Lot #** 4030410

**Made** 03/07/24



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 3

Date: 03/07/24 at 2:22 PM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**