



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 7432			
POTASSIUM PERMANGANATE 0.02 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	0.0200 EQ/L	0.0195 - 0.0205	0.0199
Sample ID			
Tested	03/12/24		
Tester	SHAR		
Recorded	03/12/24		
Traceable to N.I.S.T. SRM 723	Yes YES		YES
Sample ID			
Tested	03/12/24		
Tester	SHAR		
Recorded	03/12/24		
QCSP9:STADARDIZATION OF POTASSIUM PERMAN	YES QCSP9		YES
Sample ID			
Tested	03/12/24		
Tester	SHAR		
Recorded	03/12/24		
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	03/12/24		
Tester	SHAR		
Recorded	03/12/24		
APPEARANCE AND COLOR	CLEAR PURPLE LIQUID APPEARANCE		PASS
Sample ID			
Tested	03/12/24		
Tester	SHAR		
Recorded	03/12/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M91,VF5-L16,M55,M91,BID026 99,TID4
Sample ID			
Tested	03/12/24		
Tester	SHAR		
Recorded	03/12/24		
Actual Density	0.99865 GM/ML	0.99860 - 0.99889	0.99869
Sample ID			
Tested	03/12/24		
Tester	SHAR		
Recorded	03/12/24		
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY		03/30/2025
Sample ID			
Tested	03/12/24		
Tester	SHAR		
Recorded	03/12/24		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 03/12/24 at 4:01 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Lot #	4031156
Made	03/12/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager