



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 1792			
CHLORINE STANDARD 100 PPM AS CL2 CERTIFIED SOLUTION			
Test	Target/UOM	Range	Result
Actual Density	0.99951 GM/ML	0.99884 - 1.00018	0.99940
Sample ID Tested 03/12/24 Tester AFRE Recorded 03/12/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M55,M112,VF12-90,P167,BID0 0488M,TI
Sample ID Tested 03/12/24 Tester AFRE Recorded 03/12/24			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		05/17/24
Sample ID Tested 03/12/24 Tester AFRE Recorded 03/12/24			
SHELF LIFE OF PRODUCT	2 MONTHS TIME		2 MONTHS
Sample ID Tested 03/12/24 Tester AFRE Recorded 03/12/24			
PPM CHLORIDE	100.00 PPM	98.00 - 102.00	100.64
Sample ID Tested 03/12/24 Tester AFRE Recorded 03/12/24			
Lot #	4031103		
Made	03/12/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.