



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 7350			
POTASSIUM IODIDE-IODATE 0.1 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	0.1000	0.0995 - 0.1005	0.1000
	EQ/L		
Sample ID			
Tested	03/18/24		
Tester	SHAR		
Recorded	03/18/24		
Actual Density	1.02697	1.02568 - 1.02825	1.02731
	GM/ML		
Sample ID			
Tested	03/18/24		
Tester	SHAR		
Recorded	03/18/24		
QCSP33:POTASSIUM IODATE-IODIDE TITRANT	YES		YES
	QCSP33		
Sample ID			
Tested	03/18/24		
Tester	SHAR		
Recorded	03/18/24		
N.I.S.T. Traceable to SRM 136	Yes		YES
	YES		
Sample ID			
Tested	03/18/24		
Tester	SHAR		
Recorded	03/18/24		
PREPARED TO FORMULATION ON FILE	YES		YES
	YES		
Sample ID			
Tested	03/18/24		
Tester	SHAR		
Recorded	03/18/24		
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	03/18/24		
Tester	SHAR		
Recorded	03/18/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M103,M112,T208
Sample ID			
Tested	03/18/24		
Tester	SHAR		
Recorded	03/18/24		
SHELF LIFE OF PRODUCT	1 YEAR TIME		1 YEAR
Sample ID			
Tested	03/18/24		
Tester	SHAR		
Recorded	03/18/24		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 03/18/24 at 1:19 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

EXPIRATION DATE

MM/DD/YYYY

03/30/2025

MM/DD/YYYY

Sample ID

Tested 03/18/24

Tester SHAR

Recorded 03/18/24

Lot # 4031546

Made 03/18/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager