



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
 Customer PO # :

Packaged Product: I7615F			
ISOPROPYL 70% ALCOHOL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 03/21/24			
Tester BPHI			
Recorded 03/21/24			
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested 03/21/24			
Tester BPHI			
Recorded 03/21/24			
SPECIFIC GRAVITY @ 15.5C	0.88000	0.87200 - 0.88400	0.88160
	SPECIFIC GRAVIT		
Sample ID			
Tested 03/21/24			
Tester BPHI			
Recorded 03/21/24			
INSTRUMENTS USED DURING PREPARATION	LAB EQUIPMENT INSTRUMENT		T56
Sample ID			
Tested 03/21/24			
Tester BPHI			
Recorded 03/21/24			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		03/30/26
Sample ID			
Tested 03/21/24			
Tester BPHI			
Recorded 03/21/24			
VOLUME PERCENT ISOPROPANOL	70.0 V/V%	69.9 - 70.1	70.0
Sample ID			
Tested 03/21/24			
Tester BPHI			
Recorded 03/21/24			
SHELF LIFE OF PRODUCT	SHELF LIFE TIME		2 YEARS
Sample ID			
Tested 03/21/24			
Tester BPHI			
Recorded 03/21/24			
Actual Density	0.87943 GM/ML	0.87530 - 0.88356	0.88074
Sample ID			
Tested 03/21/24			
Tester BPHI			
Recorded 03/21/24			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 03/21/24 at 2:02 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Lot #	4032113
Made	03/21/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager