



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: CYM003F			
ACRYLAMIDE 10 PPM/ACRYLIC ACID 50 PPM			
Test	Target/UOM	Range	Result
Actual ppm Acrylamide	10.00 PPM	9.00 - 11.00	10.37
Sample ID			
Tested	03/19/24		
Tester	AROS		
Recorded	03/19/24		
Actual ppm Acrylic Acid	50.00 PPM	45.00 - 55.00	53.35
Sample ID			
Tested	03/19/24		
Tester	AROS		
Recorded	03/19/24		
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	03/19/24		
Tester	AROS		
Recorded	03/19/24		
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested	03/19/24		
Tester	AROS		
Recorded	03/19/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M108
Sample ID			
Tested	03/19/24		
Tester	AROS		
Recorded	03/19/24		
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY		05/19/2024
Sample ID			
Tested	03/19/24		
Tester	AROS		
Recorded	03/19/24		
SHELF LIFE OF PRODUCT	2 MONTHS TIME		2 MONTHS
Sample ID			
Tested	03/19/24		
Tester	AROS		
Recorded	03/19/24		
Lot #	4031855		
Made	03/19/24		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 03/19/24 at 7:34 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager