



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Test	Target/UOM	Range	Result
Packaged Product: DC517F			
	HYDROCHLORIC ACID 0.07 NORMAL		
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 04/11/24			
Tester SHAR			
Recorded 04/11/24			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 04/11/24			
Tester SHAR			
Recorded 04/11/24			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		04/11/2024
Sample ID			
Tested 04/11/24			
Tester SHAR			
Recorded 04/11/24			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	21.0
Sample ID			
Tested 04/11/24			
Tester SHAR			
Recorded 04/11/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		04/11/2024
Sample ID			
Tested 04/11/24			
Tester SHAR			
Recorded 04/11/24			
AMBIENT TEST TEMPERATURE DEG C	20.0 DEG C	15.0 - 30.0	21.0
Sample ID			
Tested 04/11/24			
Tester SHAR			
Recorded 04/11/24			
Actual Normality	0.0700 EQ/L	0.0695 - 0.0705	0.0701
Sample ID			
Tested 04/11/24			
Tester SHAR			
Recorded 04/11/24			
N.I.S.T. Traceable to SRM 723 THAM	YES YES		YES
Sample ID			
Tested 04/11/24			
Tester SHAR			
Recorded 04/11/24			




Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE	PASS
Sample ID		
Tested	04/11/24	
Tester	SHAR	
Recorded	04/11/24	
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	GC,T-3,M55,BID02699,PIDM91
Sample ID		
Tested	04/11/24	
Tester	SHAR	
Recorded	04/11/24	
Actual Density	0.99951 GM/ML	0.99944 - 0.99957 0.99954
Sample ID		
Tested	04/11/24	
Tester	SHAR	
Recorded	04/11/24	
Lot #	4041036	
Made	04/11/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



Quality Manager