



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: ND229F			
TOC 3000 MG/L & TIC 200 MG/L			
Test	Target/UOM	Range	Result
APPEARANCE AND COLOR			
	CLEAR WATER WHITE		PASS
	APPEARANCE		
Sample ID			
Tested	04/18/24		
Tester	DTRA		
Recorded	04/18/24		
CERTIFIED TOC VALUE			
	3000	2900 - 3100	2979
	PPM TOC		
Sample ID			
Tested	04/18/24		
Tester	DTRA		
Recorded	04/18/24		
CERTIFIED IC VALUE			
	200	185 - 215	189
	ACT IC		
Sample ID			
Tested	04/18/24		
Tester	DTRA		
Recorded	04/18/24		
INSTRUMENTS USED DURING PREPARATION			
	INSTRUMENT		M108,P173,VF1-Z26
Sample ID			
Tested	04/18/24		
Tester	DTRA		
Recorded	04/18/24		
EXPIRATION DATE			
	MM/DD/YY		06/23/24
	MM/DD/YYYY		
Sample ID			
Tested	04/18/24		
Tester	DTRA		
Recorded	04/18/24		
ACTUAL PPM CERTIFIED TOC VALUE			
	3000	2900 - 3100	2979
	PPM		
Sample ID			
Tested	04/18/24		
Tester	DTRA		
Recorded	04/18/24		
Lot #	4041773		
Made	04/18/24		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 04/18/24 at 2:26 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager