



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> HAC460F			
Calibration Soln 0.5 PPM CaCO3 for EZ1016-Total Hardness			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON</b>	YES		YES
<b>FILE</b>	YES		
Sample ID			
Tested 04/18/24			
Tester MMAT			
Recorded 04/18/24			
<b>CALCULATED PPM CaCO3</b>	0.50 PPM	0.47 - 0.53	0.50
Sample ID			
Tested 04/18/24			
Tester MMAT			
Recorded 04/18/24			
<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE APPEARANCE		PASS
Sample ID			
Tested 04/18/24			
Tester MMAT			
Recorded 04/18/24			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		P101,VF2-B36
Sample ID			
Tested 04/18/24			
Tester MMAT			
Recorded 04/18/24			
<b>EXPIRATION DATE</b>	MM/DD/YYYY MM/DD/YYYY		07/30/2024
Sample ID			
Tested 04/18/24			
Tester MMAT			
Recorded 04/18/24			
<b>Lot #</b>	4041809		
<b>Made</b>	04/18/24		

**Quality Manager**